



Test Report issued under the responsibility of:



Prodigy Technology Consultant Co., Ltd.

IEC / EN 60601-1
Medical electrical equipment
Part 1: General requirements for basic safety and essential performance

Report Reference No.: M170518-06-A0

Date of issue: 2017-08-22

Total number of pages: 154

Testing Laboratory.....: Prodigy Technology Consultant Co., Ltd.

Address: No.181, Sec. 2, Wunhua 1st Rd., Linkou District, New Taipei City 244, Taiwan, R.O.C.

Applicant's name: AG Neovo

Address: 5F-1, No. 3-1, Park Street, Nangang District, Taipei, 11503, Taiwan

Test specification:

Standard.....: EN 60601-1:2006 + A11: 2011 + A1: 2013 + A12: 2014;
IEC 60601-1: 2005 + CORR. 1:2006 + CORR. 2:2007 + AM1:2012 (or IEC 60601-1: 2012 reprint)

Test procedure: CB Scheme / Service in CE marking

Non-standard test method.....: None

Test Report Form No.: DTL-127-A2

Test Report Form Originator.....: Prodigy Technology Consultant Co., Ltd.

Master TRF.....: Dated 2014-07-07

Test item description: LCD Monitor

Trade Mark.....:





Manufacturer: TOP VICTORY ELECTRONICS (TAIWAN) CO LTD

Model/Type reference: MX-22x, MX-24x (where x can be A to Z (a-z), 0 to 9 or "+", "-", "\", "/")

Ratings: Power Adapter: Adapter Technology Co., Ltd. / ATM065-P120 or ATM065T-P120
Input: 100-240Vac, 50-60Hz, 1.6-0.7A;
Output: 12Vdc, 5.0A
LCD Monitor rating: 12Vdc, 2.5A



Testing procedure and testing location:	
<input checked="" type="checkbox"/> Testing Laboratory:	
Testing location/ address	Prodigy Technology Consultant Co., Ltd. / No.181, Sec. 2, Wunhua 1st Rd., Linkou District, New Taipei City 244, Taiwan, R.O.C.
<input type="checkbox"/> Associated Test Laboratory:	
Testing location/ address	
Tested by (name + signature).....	Yama Cheng / Project Handler 
Approved by (+ signature).....	Tony Hsu / Reviewer 



Certificate



Certificate no.

TA 50237927 04

License Holder:

Adapter Technology Co., Ltd.
6F-9, No. 258, Liancheng Rd.,
Zhonghe District, New Taipei City
235
Taiwan

Manufacturing Plant:

Boayang Electronics Co., Ltd.
Di Feng Gong Ye Qu 2 Hao,
Xiasha Liuwu Village,
Shipai Town, Dong Guan City,
Guang Dong Province 523350
P.R. China

Test report no.: ZTWI- 10037235 005

Client Reference: TH/SPC-1611073-Px

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Medical Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

Remark: 1. The usability requirements of the standard were not addressed.
2. This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Date of Issue

(day/mo/yr)

25/01/2017



Dipl.-Ing. (FH) A. Klinker



Certificate



Certificate no.

TA 50368835 01

License Holder:

Adapter Technology Co., Ltd.
6F-9, No. 258, Liancheng Rd.,
Zhonghe District, New Taipei City
235
Taiwan

Manufacturing Plant:

Boayang Electronics Co., Ltd.
Di Feng Gong Ye Qu 2 Hao,
Xiasha Liuwu Village,
Shipai Town, Dong Guan City,
Guang Dong Province 523350
P.R. China

Test report no.: ZTW1-LinPr 50059724 001 Client Reference: TH/SPC-1608077

Tested to:
IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-6:2010+A1
EN 60601-1-6:2010+A1
IEC 62366:2007+A1
EN 62366:2008+A1

Certified Product: (Medical Power Supply) **License Fee - Units**

Type Designation	: ATM065T-Px (ADAPTER TECH.)	8
x stands for	: 120, 150, 180, 190, 240 or 480	1
Rated Voltage	: AC 100-240V, 50-60Hz	
Rated Current	: 1.6-0.7A	
Output Voltage	: see constructional dataform	
Output Current	: see constructional dataform	
max. Ambient Temperature	: +40°C	
Protection Class	: I	

This certificate is valid for maximum 5 years.

9

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. (FH) A. Klinker



Date of Issue
(day/mo/yr)

13/01/2017